

**Shenzhen CTO Technology Service Co.,Ltd**

9 / F, Block B, 196 Tangtou Avenue, Langxin Community,  
Shiyan Street, Baoan District, Shenzhen China.



## CERTIFICATE OF CONFORMITY

**Certificate No.** : CTO200318074ECS

**Product** : INFRARED BODY THERMOMETER

**Trade Name** : *AFK*

**Model(s)** : YK001, YK002, YK003, YK004

**Test Report No.** : CTO200312040DRS  
CTO200312012URE  
EN 60601-1: 2006 + A1: 2013+A12:2014

**Test Standards** : EN 60601-1-11:2015  
EN 60601-1-2:2015

The EUT described above has been tested by us with the listed standards and found in compliance with the council **Medical Devices Directive(MDD) 93/42/EEC**. It is possible to use **CE** marking to demonstrate the compliance with this **MD Directive**.

**CE**

  
For Chief Executive / Zhenhua Lee  
Date: March 18, 2020

This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole production and other relevant directives have to be observed.

Web: [www.cto-cert.com](http://www.cto-cert.com)

E-mail: [service@cto-cert.com](mailto:service@cto-cert.com)